
Midmark M11 - v1.0.5
Total Cycles: 1022

Name: RITTER2
Sterilizer ID: V2313228
04/22/2022 01:18 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	76.9 F	0.0
2:00	80.9 F	0.0
4:00	110.2 F	0.0
6:00	127.6 F	0.1
8:00	162.1 F	0.6
10:00	182.9 F	4.7
12:00	201.6 F	9.2
14:00	211.8 F	10.3
16:00	226.1 F	7.5
18:00	237.5 F	12.1
20:00	236.5 F	9.4
22:00	239.5 F	9.7
24:00	248.2 F	14.3
26:00	254.3 F	17.5

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 247F
CHAMBER PRESS 13.6PSI
** ITEMS NOT STERILE **

04/22/2022 01:46 PM