
Midmark M11 - v1.0.5
Total Cycles: 2682

Name: RITTER2
Sterilizer ID: V2313228
07/21/2023 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.4 F 0.0
2:00 145.7 F 0.0
4:00 152.6 F 1.0
6:00 182.0 F 4.3
8:00 200.0 F 8.0
10:00 209.5 F 12.2
12:00 221.5 F 6.7
14:00 231.8 F 11.2
16:00 232.7 F 8.1
18:00 242.8 F 12.8
20:00 243.1 F 12.6
22:00 254.6 F 18.6
24:00 265.2 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 30.0
2:00 271.6 F 29.4
3:00 271.7 F 29.9
4:00 272.6 F 29.9

Min 271.2 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 211.1 F 0.0
10:00 222.3 F 0.0
15:00 218.1 F 0.0
20:00 210.0 F 0.0
25:00 214.3 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:55
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

07/21/2023 05:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____