
Midmark M11 - v1.0.5
Total Cycles: 684

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 77.2 F 0.0
4:00 106.0 F 0.0
6:00 114.6 F 0.2
8:00 144.0 F 2.3
10:00 168.5 F 5.2
12:00 184.2 F 8.8
14:00 201.0 F 12.7
16:00 215.0 F 6.3
18:00 227.6 F 10.5
20:00 230.5 F 8.5
22:00 240.0 F 12.4
24:00 239.7 F 11.0
26:00 252.4 F 17.3
28:00 263.2 F 23.7
30:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 30.1
2:00 271.8 F 29.3
3:00 272.2 F 30.0
4:00 272.4 F 29.8

Min 271.3 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 197.3 F 0.0
10:00 223.1 F 0.0
15:00 219.7 F 0.0
20:00 212.8 F 0.0
25:00 226.7 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:14
CYCLE COMPLETE

01/12/2022 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____