

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000524  
14:20 28/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:12  
STERILIZING 4:23  
136.4 C 231kPa 4:23  
136.2 C 232kPa 4:53  
136.4 C 233kPa 5:23  
136.3 C 233kPa 5:53  
136.4 C 233kPa 6:23  
136.2 C 231kPa 6:53  
136.3 C 232kPa 7:23  
136.4 C 233kPa 7:53  
136.4 C 233kPa 7:53  
VENTING 7:53  
AIR DRYING 8:48  
CYCLE COMPLETE 23:48  
-----

Digital Signature #  
7EAE852E88945A24  
-----