
Midmark M11 - v1.0.5
Total Cycles: 802

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.5 F 0.0
2:00 175.4 F 3.4
4:00 194.4 F 8.1
6:00 216.4 F 12.6
8:00 233.5 F 9.2
10:00 235.3 F 9.5
12:00 245.3 F 13.2
14:00 244.6 F 12.5
16:00 257.8 F 19.6
18:00 269.4 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.1
1:00 272.3 F 29.3
2:00 272.2 F 29.3
3:00 272.1 F 29.2
4:00 272.2 F 29.1

Min 271.9 F 29.0
Max 273.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 231.2 F 0.0
10:00 234.2 F 0.0
15:00 227.1 F 0.0
20:00 228.2 F 0.0
25:00 227.0 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

02/15/2022 01:26 PM
ACCEPTED: X
REJECTED:
Operator: U02