
Midmark M11 - v1.0.0
Total Cycles: 2778

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.8 F 0.0
2:00 205.1 F 1.6
4:00 221.3 F 7.2
6:00 233.8 F 12.7
8:00 234.0 F 9.1
10:00 240.3 F 9.7
12:00 246.2 F 13.4
14:00 245.2 F 12.6
16:00 258.3 F 19.7
18:00 269.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.8
2:00 271.5 F 28.1
3:00 272.5 F 28.9
4:00 271.9 F 28.4

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 223.1 F 0.0
10:00 238.8 F 0.0
15:00 235.6 F 0.0
20:00 227.4 F 0.0
25:00 225.4 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

01/12/2023 05:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____