
Midmark M11 - v1.0.5
Total Cycles: 4137

Name: RITTER2
Sterilizer ID: V2313228
11/19/2024 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.5 F 0.0
2:00 131.2 F 0.1
4:00 149.3 F 1.2
6:00 178.5 F 5.0
8:00 203.4 F 9.0
10:00 219.5 F 11.4
12:00 226.8 F 8.2
14:00 236.3 F 13.0
16:00 237.4 F 10.5
18:00 235.6 F 8.9
20:00 249.9 F 15.6
22:00 261.2 F 22.1
24:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.8 F 29.7
2:00 271.9 F 29.4
3:00 272.2 F 29.3
4:00 272.9 F 29.6

Min 271.0 F 28.8
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 232.3 F 0.0
10:00 226.6 F 0.0
15:00 186.0 F 0.0
20:00 177.3 F 0.0
25:00 172.7 F 0.0
30:00 170.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:37
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:03:06
CYCLE COMPLETE

11/19/2024 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____