
Midmark M11 - v1.0.0
Total Cycles: 4930

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 146.9 F 0.0
4:00 146.5 F 0.1
6:00 182.5 F 0.3
8:00 203.2 F 1.6
10:00 212.7 F 5.5
12:00 223.6 F 9.1
14:00 234.5 F 13.2
16:00 235.4 F 9.5
18:00 240.0 F 10.1
20:00 244.7 F 12.8
22:00 242.8 F 11.5
24:00 254.9 F 17.7
26:00 265.4 F 24.0
28:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.7
2:00 272.6 F 28.8
3:00 272.1 F 28.5
4:00 271.6 F 28.1

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 209.2 F 0.0
10:00 209.6 F 0.0
15:00 208.8 F 0.0
20:00 209.4 F 0.0
25:00 212.6 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 28:03
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:31
CYCLE COMPLETE

09/23/2024 11:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____