
Midmark M11 - v1.0.5
Total Cycles: 2271

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.5 F 0.0
2:00 190.7 F 4.2
4:00 213.7 F 10.0
6:00 225.1 F 6.8
8:00 239.4 F 12.6
10:00 239.5 F 11.1
12:00 239.5 F 10.4
14:00 253.9 F 17.6
16:00 266.0 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.1 F 28.9
2:00 272.3 F 29.4
3:00 271.4 F 28.9
4:00 272.3 F 29.1

Min 271.3 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 222.9 F 0.0
10:00 233.1 F 0.0
15:00 226.7 F 0.0
20:00 229.7 F 0.0
25:00 227.2 F 0.0
30:00 225.6 F 0.0

Drying Complete

FILLING: 2:47
HEATING: 17:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

04/19/2023 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____