
Midmark M11 - v1.0.5
Total Cycles: 1867

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 05:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 176.9 F 4.1
4:00 190.3 F 8.7
6:00 210.1 F 8.5
8:00 227.7 F 9.0
10:00 235.5 F 9.7
12:00 242.7 F 12.9
14:00 243.0 F 12.1
16:00 256.5 F 19.2
18:00 268.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.6
2:00 271.5 F 29.0
3:00 272.7 F 29.7
4:00 272.1 F 29.1

Min 271.5 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 228.1 F 0.0
10:00 225.6 F 0.0
15:00 224.5 F 0.0
20:00 225.2 F 0.0
25:00 225.4 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

12/20/2022 06:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____