
Midmark M11 - v1.0.0
Total Cycles: 1823

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 07:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 85.4 F 0.0
4:00 101.2 F 0.0
6:00 129.2 F 0.1
8:00 160.9 F 1.0
10:00 182.0 F 4.6
12:00 201.9 F 8.4
14:00 218.6 F 12.1
16:00 222.4 F 5.5
18:00 233.4 F 9.7
20:00 235.8 F 9.8
22:00 242.1 F 11.7
24:00 239.2 F 10.0
26:00 252.9 F 16.7
28:00 264.3 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.6 F 28.9
2:00 271.9 F 28.5
3:00 271.9 F 28.3
4:00 272.6 F 29.0

Min 271.6 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 227.9 F 0.0
10:00 207.4 F 0.0
15:00 204.8 F 0.0
20:00 211.3 F 0.0
25:00 220.0 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:46
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:37
CYCLE COMPLETE

04/14/2022 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____