
Midmark M11 - v1.0.5
Total Cycles: 570

Name: RITTER2
Sterilizer ID: V2313228
11/23/2021 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 140.4 F 1.6
4:00 155.4 F 5.0
6:00 182.9 F 9.7
8:00 207.6 F 5.1
10:00 218.3 F 8.4
12:00 233.6 F 11.1
14:00 238.3 F 11.2
16:00 237.3 F 9.6
18:00 251.1 F 16.4
20:00 262.9 F 23.2
22:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.1
1:00 272.4 F 29.8
2:00 272.0 F 29.0
3:00 272.4 F 29.9
4:00 272.2 F 29.2

Min 271.4 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 233.8 F 0.0
10:00 227.9 F 0.0
15:00 227.9 F 0.0
20:00 224.7 F 0.0
25:00 224.3 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:04
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

11/23/2021 03:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____