

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4654

Name: RITTER 1  
Sterilizer ID: V2178198  
06/18/2024 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.3 F 0.0  
2:00 89.4 F 0.2  
4:00 115.5 F 0.9  
6:00 145.1 F 3.5  
8:00 171.8 F 6.7  
10:00 191.3 F 10.1  
12:00 207.4 F 9.8  
14:00 216.9 F 6.7  
16:00 229.7 F 11.0  
18:00 229.1 F 6.6  
20:00 241.4 F 12.2  
22:00 239.0 F 10.0  
24:00 251.6 F 16.2  
26:00 262.7 F 22.6  
28:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.9  
2:00 271.5 F 28.5  
3:00 272.6 F 28.8  
4:00 271.5 F 28.4

Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 214.1 F 0.0  
10:00 221.2 F 0.0  
15:00 226.9 F 0.0  
20:00 227.8 F 0.0  
25:00 229.8 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:18  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:06:36  
CYCLE COMPLETE

06/18/2024 03:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_