

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2552

Name: RITTER 1  
Sterilizer ID: V2178198  
11/14/2022 01:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 186.4 F 4.2  
4:00 190.6 F 8.7  
6:00 210.1 F 10.1  
8:00 223.6 F 7.8  
10:00 234.8 F 12.9  
12:00 236.9 F 9.9  
14:00 237.6 F 8.4  
16:00 249.2 F 14.8  
18:00 261.3 F 21.6  
20:00 272.3 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.1  
2:00 271.8 F 28.4  
3:00 272.0 F 28.9  
4:00 272.4 F 28.9

Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 229.5 F 0.0  
10:00 238.2 F 0.0  
15:00 231.0 F 0.0  
20:00 229.1 F 0.0  
25:00 229.8 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

11/14/2022 02:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_