

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3467

Name: RITTER2  
Sterilizer ID: V2313228  
03/21/2024 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.8 F 0.0  
2:00 174.2 F 3.0  
4:00 193.2 F 7.6  
6:00 211.5 F 12.4  
8:00 223.1 F 7.3  
10:00 232.6 F 12.6  
12:00 238.5 F 10.5  
14:00 236.9 F 9.5  
16:00 251.2 F 16.2  
18:00 262.7 F 22.9  
20:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.4 F 29.5  
2:00 271.6 F 28.9  
3:00 271.4 F 29.3  
4:00 273.1 F 30.1

Min 271.0 F 28.7  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 232.6 F 0.0  
10:00 225.0 F 0.0  
15:00 219.4 F 0.0  
20:00 216.6 F 0.0  
25:00 218.9 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

03/21/2024 04:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_