
Midmark M11 - v1.0.0
Total Cycles: 5222

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2025 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.0 F 0.0
2:00 111.2 F 0.0
4:00 146.0 F 0.0
6:00 165.5 F 0.5
8:00 190.2 F 2.7
10:00 201.7 F 6.1
12:00 212.9 F 9.5
14:00 224.2 F 13.4
16:00 229.1 F 7.8
18:00 239.7 F 12.4
20:00 237.6 F 9.6
22:00 239.5 F 9.7
24:00 247.8 F 13.9
26:00 259.1 F 20.0
28:00 269.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.7
2:00 271.9 F 28.1
3:00 271.8 F 28.3
4:00 272.6 F 28.6

Min 271.4 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 200.8 F 0.0
10:00 199.6 F 0.0
15:00 199.3 F 0.0
20:00 198.9 F 0.0
25:00 199.9 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:06
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:36
CYCLE COMPLETE

01/06/2025 04:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____