

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1514

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2022 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 189.0 F 3.4  
4:00 185.7 F 7.5  
6:00 207.2 F 13.1  
8:00 221.4 F 7.2  
10:00 235.6 F 12.3  
12:00 236.1 F 9.5  
14:00 238.5 F 9.2  
16:00 247.3 F 14.0  
18:00 259.3 F 20.6  
20:00 270.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.7 F 29.2  
2:00 271.6 F 28.5  
3:00 272.7 F 29.1  
4:00 271.6 F 28.5  
  
Min 271.5 F 28.4  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 234.4 F 0.0  
10:00 227.7 F 0.0  
15:00 230.2 F 0.0  
20:00 229.1 F 0.0  
25:00 230.8 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

01/12/2022 12:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_