
Midmark M11 - v1.0.5
Total Cycles: 21

Name: RITTER2
Sterilizer ID: V2313228
04/26/2021 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.8 F 0.0
2:00 179.5 F 4.4
4:00 202.6 F 10.0
6:00 224.5 F 6.3
8:00 236.3 F 12.6
10:00 239.0 F 11.2
12:00 239.7 F 10.7
14:00 254.1 F 17.9
16:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.4
2:00 271.6 F 29.2
3:00 272.4 F 29.4
4:00 271.5 F 28.9

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 225.5 F 0.0
10:00 225.8 F 0.0
15:00 219.5 F 0.0
20:00 216.4 F 0.0
Dry Cycle Aborted

FILLING: 2:57
HEATING: 17:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 20:46
TOTAL CYCLE: 00:48:26

04/26/2021 03:30 PM
ACCEPTED: X
REJECTED:
Operator: U01