
Midmark M11 - v1.0.5
Total Cycles: 3305

Name: RITTER2
Sterilizer ID: V2313228
01/17/2024 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 77.6 F 0.0
4:00 75.2 F 0.1
6:00 100.3 F 1.1
8:00 140.9 F 3.8
10:00 168.5 F 6.7
12:00 189.3 F 10.1
14:00 215.6 F 7.6
16:00 219.2 F 7.3
18:00 231.2 F 11.5
20:00 229.3 F 7.6
22:00 242.3 F 13.1
24:00 240.9 F 11.5
26:00 253.3 F 17.7
28:00 264.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 271.6 F 29.6
2:00 272.8 F 29.7
3:00 271.1 F 29.0
4:00 272.1 F 29.9

Min 271.0 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 219.1 F 0.0
10:00 186.4 F 0.0
15:00 186.8 F 0.0
20:00 185.6 F 0.0
25:00 185.6 F 0.0
30:00 187.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:01
CYCLE COMPLETE

01/17/2024 10:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____