
Midmark M11 - v1.0.0
Total Cycles: 2159

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2022 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 75.0 F 0.0
4:00 81.8 F 0.0
6:00 107.7 F 0.2
8:00 143.7 F 1.9
10:00 167.4 F 5.1
12:00 204.7 F 8.5
14:00 214.0 F 12.1
16:00 215.1 F 4.9
18:00 226.4 F 8.7
20:00 234.9 F 12.7
22:00 234.4 F 9.0
24:00 241.0 F 11.2
26:00 243.4 F 12.2
28:00 255.1 F 18.2
30:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 29.2
2:00 272.2 F 28.6
3:00 271.2 F 28.6
4:00 272.5 F 29.2

Min 271.0 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 188.9 F 0.0
10:00 194.5 F 0.0
15:00 178.4 F 0.0
20:00 171.2 F 0.0
25:00 170.4 F 0.0
30:00 171.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

07/21/2022 08:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____