
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002462
11:20 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:17
STERILIZING 4:23
136.0 C 221kPa 4:23
136.0 C 223kPa 4:53
136.3 C 225kPa 5:23
136.4 C 225kPa 5:53
136.3 C 224kPa 6:23
136.4 C 226kPa 6:53
136.4 C 226kPa 7:23
136.4 C 224kPa 7:53
136.4 C 224kPa 7:53
VENTING 7:53
AIR DRYING 8:48
CYCLE COMPLETE 9:01

Digital Signature #
BE7F05970C703940
