
Midmark M11 - v1.0.5
Total Cycles: 812

Name: RITTER2
Sterilizer ID: V2313228
02/17/2022 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 75.2 F 0.0
4:00 90.3 F 0.0
6:00 124.8 F 0.3
8:00 153.0 F 2.6
10:00 178.7 F 6.0
12:00 196.4 F 9.7
14:00 208.4 F 10.0
16:00 220.6 F 7.4
18:00 232.5 F 11.6
20:00 232.4 F 8.2
22:00 243.3 F 12.7
24:00 242.8 F 12.2
26:00 254.6 F 18.3
28:00 265.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 30.1
2:00 272.0 F 29.9
3:00 271.6 F 29.5
4:00 272.2 F 29.6

Min 270.6 F 28.8
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 183.3 F 0.0
10:00 188.7 F 0.0
15:00 208.7 F 0.0
20:00 209.8 F 0.0
25:00 219.5 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 29:39
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:53
CYCLE COMPLETE

02/17/2022 09:31 AM
ACCEPTED: X
REJECTED:
Operator: U02