

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3865

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 05:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
2:00 174.7 F 3.3  
4:00 189.1 F 7.0  
6:00 208.0 F 11.6  
8:00 221.7 F 5.3  
10:00 233.4 F 9.7  
12:00 236.0 F 9.2  
14:00 243.3 F 12.3  
16:00 241.6 F 11.0  
18:00 254.5 F 17.5  
20:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.2  
1:00 272.2 F 28.7  
2:00 272.2 F 28.9  
3:00 271.6 F 28.2  
4:00 272.1 F 28.7  
  
Min 271.4 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 224.2 F 0.0  
10:00 196.2 F 0.0  
15:00 214.8 F 0.0  
20:00 211.6 F 0.0  
25:00 210.8 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

10/04/2023 06:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_