

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001443  
11:27 07/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:09  
PRESSURIZING 2:15  
STERILIZING 3:13  
136.4 C 220kPa 3:13  
136.4 C 226kPa 3:43  
136.3 C 223kPa 4:13  
136.4 C 226kPa 4:43  
136.3 C 225kPa 5:13  
136.1 C 223kPa 5:43  
136.4 C 225kPa 6:13  
136.3 C 225kPa 6:43  
136.3 C 225kPa 6:43  
VENTING 6:43  
AIR DRYING 7:35  
CYCLE COMPLETE 22:35  
-----

Digital Signature #  
58A8E98FEC561608  
-----