
Midmark M11 - v1.0.0
Total Cycles: 4459

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.8 F 0.0
2:00 124.5 F 1.4
4:00 143.6 F 4.1
6:00 170.5 F 7.2
8:00 204.9 F 11.1
10:00 220.0 F 3.7
12:00 219.3 F 7.5
14:00 229.4 F 11.5
16:00 231.1 F 7.2
18:00 242.3 F 12.5
20:00 239.9 F 10.5
22:00 251.9 F 16.2
24:00 262.1 F 22.2
26:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.0
2:00 272.2 F 29.3
3:00 271.6 F 28.8
4:00 272.4 F 29.4

Min 271.1 F 28.6
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 232.2 F 0.0
10:00 221.5 F 0.0
15:00 185.0 F 0.0
20:00 176.5 F 0.0
25:00 174.1 F 0.0
30:00 172.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 27:16
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:07
CYCLE COMPLETE

04/15/2024 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____