
Midmark M11 - v1.0.5
Total Cycles: 892

Name: RITTER2
Sterilizer ID: V2313228
03/21/2022 03:11 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.6 F 0.0
2:00 181.2 F 0.2
4:00 190.3 F 3.5
6:00 214.5 F 9.7
8:00 227.1 F 6.7
10:00 242.8 F 13.0
12:00 243.9 F 12.8
14:00 244.3 F 12.6
16:00 258.6 F 20.3

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 255F
CHAMBER PRESS 14.1PSI
** ITEMS NOT STERILE **

03/21/2022 03:30 PM