

-----  
Midmark M11 - v1.0.5  
Total Cycles: 203

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2021 06:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.8 F 0.0  
2:00 72.4 F 0.0  
4:00 74.7 F 0.2  
6:00 90.6 F 1.6  
8:00 122.6 F 3.9  
10:00 150.3 F 6.7  
12:00 170.9 F 10.0  
14:00 191.0 F 13.4  
16:00 203.0 F 5.5  
18:00 214.8 F 8.9  
20:00 226.9 F 12.9  
22:00 229.3 F 9.8  
24:00 234.9 F 9.4  
26:00 246.4 F 13.8  
28:00 257.6 F 19.9  
30:00 265.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.8  
1:00 272.3 F 29.9  
2:00 271.6 F 30.1  
3:00 271.9 F 29.8  
4:00 271.1 F 30.2

Min 270.9 F 29.7  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 230.1 F 0.0  
10:00 210.7 F 0.0  
15:00 207.1 F 0.0  
20:00 207.6 F 0.0  
25:00 211.5 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 31:21  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:09:25  
CYCLE COMPLETE

07/14/2021 07:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01