
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000695
14:06 22/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:35
STERILIZING 3:30
136.5 C 222kPa 3:30
136.4 C 225kPa 4:00
136.4 C 223kPa 4:30
136.3 C 224kPa 5:00
136.3 C 222kPa 5:30
136.3 C 222kPa 6:00
136.4 C 223kPa 6:30
135.8 C 220kPa 7:00
135.8 C 220kPa 7:00
VENTING 7:00
AIR DRYING 7:49
CYCLE COMPLETE 7:54

Digital Signature #
96C0C0CACCE28124
