
Midmark M11 - v1.0.0
Total Cycles: 4026

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 06:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.2 F 0.0
2:00 139.1 F 1.2
4:00 133.1 F 3.8
6:00 155.5 F 7.1
8:00 184.5 F 10.8
10:00 215.6 F 4.0
12:00 221.1 F 7.5
14:00 231.2 F 11.7
16:00 230.4 F 7.7
18:00 242.6 F 12.9
20:00 240.8 F 10.9
22:00 252.3 F 16.8
24:00 262.4 F 22.9
26:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 273.0 F 29.6
2:00 272.2 F 29.1
3:00 271.3 F 28.8
4:00 272.6 F 29.7

Min 271.3 F 28.7
Max 273.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 231.2 F 0.0
10:00 226.5 F 0.0
15:00 229.0 F 0.0
20:00 228.9 F 0.0
25:00 225.2 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 26:29
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:18
CYCLE COMPLETE

11/15/2023 07:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____