
Midmark M11 - v1.0.0
Total Cycles: 5162

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2024 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 174.9 F 0.2
4:00 186.7 F 0.6
6:00 208.4 F 1.3
8:00 221.6 F 5.4
10:00 232.9 F 9.6
12:00 235.8 F 9.8
14:00 241.7 F 11.8
16:00 239.2 F 10.2
18:00 236.9 F 9.1
20:00 252.0 F 16.3
22:00 264.0 F 23.3
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 29.0
2:00 272.6 F 28.8
3:00 272.0 F 28.4
4:00 271.5 F 28.3

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 216.6 F 0.0
10:00 255.4 F 0.0
15:00 243.7 F 0.0
20:00 228.4 F 0.0
25:00 229.4 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

12/05/2024 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____