
Midmark M11 - v1.0.5
Total Cycles: 351

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 179.9 F 3.2
4:00 197.4 F 8.6
6:00 218.1 F 8.6
8:00 232.5 F 10.3
10:00 232.6 F 8.1
12:00 239.9 F 10.7
14:00 247.7 F 14.7
16:00 260.6 F 21.9
18:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.0 F 29.0
2:00 272.2 F 29.0
3:00 272.5 F 29.3
4:00 273.0 F 29.7

Min 271.6 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 237.2 F 0.0
10:00 224.7 F 0.0
15:00 223.1 F 0.0
20:00 217.8 F 0.0
25:00 214.4 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

09/09/2021 05:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____