

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1301

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2021 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
2:00 174.8 F 1.9  
4:00 180.2 F 4.9  
6:00 202.7 F 9.4  
8:00 216.6 F 7.3  
10:00 227.5 F 8.5  
12:00 240.0 F 13.4  
14:00 239.1 F 10.8  
16:00 236.8 F 8.9  
18:00 250.6 F 15.7  
20:00 262.1 F 22.2  
22:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 29.3  
2:00 272.0 F 28.6  
3:00 271.6 F 28.6  
4:00 272.7 F 29.3

Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 220.4 F 0.0  
10:00 227.8 F 0.0  
15:00 226.7 F 0.0  
20:00 226.7 F 0.0  
25:00 226.6 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 22:27  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:01:16  
CYCLE COMPLETE

10/26/2021 11:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01