
Midmark M11 - v1.0.0
Total Cycles: 985

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 189.7 F 3.7
4:00 211.1 F 9.0
6:00 226.0 F 8.0
8:00 234.8 F 10.2
10:00 233.6 F 7.7
12:00 244.4 F 12.3
14:00 245.7 F 12.9
16:00 259.1 F 20.0
18:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.0 F 28.2
2:00 272.6 F 28.7
3:00 271.9 F 28.6
4:00 271.7 F 28.2

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 200.7 F 0.0
10:00 208.9 F 0.0
15:00 207.5 F 0.0
20:00 204.7 F 0.0
25:00 201.2 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 2:52
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/08/2021 01:18 PM
ACCEPTED: X
REJECTED:
Operator: U02