

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2969

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2023 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 182.1 F 3.4  
4:00 202.1 F 8.6  
6:00 218.1 F 11.0  
8:00 233.8 F 9.5  
10:00 236.5 F 9.5  
12:00 244.4 F 13.2  
14:00 245.0 F 13.1  
16:00 258.0 F 19.9  
18:00 269.2 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 30.1  
2:00 272.4 F 29.5  
3:00 271.5 F 29.1  
4:00 272.3 F 29.9  
  
Min 271.4 F 29.1  
Max 273.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 231.7 F 0.0  
10:00 190.5 F 0.0  
15:00 189.2 F 0.0  
20:00 190.8 F 0.0  
25:00 186.9 F 0.0  
30:00 185.7 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 19:52  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 00:59:33  
CYCLE COMPLETE

10/04/2023 03:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_