
Midmark M11 - v1.0.0
Total Cycles: 813

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 171.0 F 2.4
4:00 177.7 F 6.2
6:00 200.1 F 10.8
8:00 217.2 F 5.1
10:00 227.3 F 10.6
12:00 230.7 F 7.1
14:00 243.4 F 12.9
16:00 243.4 F 12.2
18:00 256.1 F 18.8
20:00 267.2 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.3 F 29.4
2:00 272.3 F 29.0
3:00 271.4 F 28.4
4:00 271.6 F 29.0

Min 271.0 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 207.7 F 0.0
10:00 218.7 F 0.0
15:00 211.6 F 0.0
20:00 233.4 F 0.0
25:00 234.1 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

04/27/2021 11:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____