
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001120
9:21 18/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:43
136.4 C 223kPa 4:43
136.0 C 217kPa 5:13
136.4 C 224kPa 5:43
136.0 C 220kPa 6:13
136.3 C 224kPa 6:43
136.2 C 222kPa 7:13
136.4 C 223kPa 7:43
136.3 C 220kPa 8:13
136.3 C 220kPa 8:13
VENTING 8:13
AIR DRYING 9:08
CYCLE COMPLETE 9:10

Digital Signature #
5022C5912B50BB59
