
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001292
15:57 12/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:48
PRESSURIZING 4:48
STERILIZING 6:03
136.0 C 221kPa 6:03
136.4 C 222kPa 6:33
136.3 C 223kPa 7:03
136.2 C 223kPa 7:33
136.4 C 225kPa 8:03
136.4 C 225kPa 8:33
136.4 C 225kPa 9:03
136.4 C 223kPa 9:33
136.4 C 223kPa 9:33
VENTING 9:33
AIR DRYING 10:32
CYCLE COMPLETE 21:44

Digital Signature #
4C09D3BD6D778EE6
