

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3354

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2023 03:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
2:00 195.3 F 2.2  
4:00 217.5 F 8.0  
6:00 226.4 F 11.3  
8:00 233.0 F 9.9  
10:00 236.6 F 8.4  
12:00 242.5 F 11.3  
14:00 246.6 F 13.6  
16:00 259.4 F 20.7  
18:00 270.8 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.4 F 29.2  
2:00 271.6 F 28.4  
3:00 272.6 F 29.2  
4:00 271.8 F 28.6  
  
Min 271.4 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.7 F 0.0  
5:00 221.4 F 0.0  
10:00 234.4 F 0.0  
15:00 228.7 F 0.0  
20:00 228.6 F 0.0  
25:00 229.4 F 0.0  
30:00 229.3 F 0.0  
  
Drying Complete

FILLING: 1:43  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

06/07/2023 04:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_