

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003837  
8:25 20/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:14  
STERILIZING 4:19  
136.3 C 220kPa 4:19  
136.3 C 223kPa 4:49  
136.4 C 221kPa 5:19  
135.9 C 221kPa 5:49  
136.4 C 222kPa 6:19  
136.4 C 224kPa 6:49  
136.4 C 224kPa 7:19  
136.4 C 221kPa 7:49  
136.4 C 221kPa 7:49  
VENTING 7:49  
AIR DRYING 8:37  
CYCLE COMPLETE 8:52

-----  
Digital Signature #  
A7F2476EF1FCB102  
-----