

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4909

Name: RITTER 1  
Sterilizer ID: V2178198  
09/13/2024 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.4 F 0.0  
2:00 149.7 F 0.2  
4:00 178.2 F 0.5  
6:00 205.0 F 2.9  
8:00 221.7 F 6.9  
10:00 235.6 F 11.0  
12:00 227.1 F 5.9  
14:00 239.5 F 11.1  
16:00 235.0 F 8.2  
18:00 242.1 F 10.7  
20:00 247.0 F 13.4  
22:00 259.2 F 20.0  
24:00 270.1 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.8 F 28.6  
2:00 271.8 F 27.8  
3:00 272.1 F 28.4  
4:00 272.6 F 28.5

Min 271.5 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.6 F 0.0  
5:00 205.6 F 0.0  
10:00 205.5 F 0.0  
15:00 244.1 F 0.0  
20:00 231.0 F 0.0  
25:00 233.3 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:03:13  
CYCLE COMPLETE

09/13/2024 03:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_