
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000554
14:23 23/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:12
STERILIZING 3:20
136.0 C 220kPa 3:20
136.5 C 233kPa 3:50
136.3 C 231kPa 4:20
136.3 C 232kPa 4:50
136.1 C 232kPa 5:20
136.4 C 233kPa 5:50
136.2 C 232kPa 6:20
136.3 C 230kPa 6:50
136.3 C 230kPa 6:50
VENTING 6:50
AIR DRYING 7:41
CYCLE COMPLETE 22:42

Digital Signature #
4A235F7D3F8D7F0B
