

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3732

Name: RITTER2  
Sterilizer ID: V2313228  
06/25/2024 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 71.1 F 0.1  
4:00 82.8 F 0.3  
6:00 104.2 F 1.8  
8:00 140.4 F 4.9  
10:00 165.8 F 8.3  
12:00 190.7 F 12.1  
14:00 203.7 F 3.7  
16:00 215.2 F 7.0  
18:00 226.6 F 10.8  
20:00 229.1 F 7.7  
22:00 238.9 F 11.7  
24:00 237.2 F 9.7  
26:00 250.0 F 15.9  
28:00 260.8 F 22.1  
30:00 270.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 273.0 F 29.7  
2:00 271.7 F 29.1  
3:00 273.1 F 29.8  
4:00 271.9 F 29.0

Min 271.6 F 28.9  
Max 273.2 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.6 F 0.0  
5:00 194.4 F 0.0  
10:00 196.6 F 0.0  
15:00 196.5 F 0.0  
20:00 196.7 F 0.0  
25:00 197.6 F 0.0  
30:00 197.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:46  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:08:58  
CYCLE COMPLETE

06/25/2024 10:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_