
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005889
12:08 04/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:07
STERILIZING 4:01
136.4 C 223kPa 4:01
136.2 C 224kPa 4:31
136.3 C 223kPa 5:01
136.4 C 224kPa 5:31
136.4 C 225kPa 6:01
136.2 C 223kPa 6:31
136.4 C 225kPa 7:01
136.4 C 223kPa 7:31
136.4 C 223kPa 7:31
VENTING 7:32
AIR DRYING 8:28
CASSETTE REMOVED
CYCLE COMPLETE 30:53

Digital Signature #
C41CDF02AE36DE41
