

-----  
Midmark M11 - v1.0.5  
Total Cycles: 322

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2021 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.8 F 0.0  
2:00 128.8 F 1.2  
4:00 145.3 F 3.8  
6:00 172.6 F 7.4  
8:00 192.3 F 11.2  
10:00 218.2 F 4.9  
12:00 233.6 F 10.3  
14:00 232.9 F 8.1  
16:00 242.0 F 12.8  
18:00 241.1 F 11.5  
20:00 253.8 F 17.9  
22:00 264.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.9 F 29.9  
2:00 271.2 F 29.1  
3:00 272.3 F 30.2  
4:00 273.2 F 30.0

Min 271.0 F 29.1  
Max 273.6 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
5:00 237.4 F 0.0  
10:00 223.0 F 0.0  
15:00 230.5 F 0.0  
20:00 226.1 F 0.0  
25:00 231.9 F 0.0  
30:00 230.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:58  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

08/30/2021 03:07 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02