
Midmark M11 - v1.0.5
Total Cycles: 801

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 189.3 F 4.2
4:00 211.2 F 10.2
6:00 228.3 F 6.6
8:00 238.8 F 13.3
10:00 241.1 F 11.9
12:00 242.1 F 11.5
14:00 256.2 F 18.9
16:00 268.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.1 F 29.0
2:00 272.4 F 29.6
3:00 271.7 F 28.8
4:00 272.5 F 29.5

Min 271.2 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 227.3 F 0.0
10:00 214.5 F 0.0
15:00 229.1 F 0.0
20:00 213.9 F 0.0
25:00 227.9 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 3:20
HEATING: 17:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

02/15/2022 12:07 PM
ACCEPTED: X
REJECTED:
Operator: U02