
Midmark M11 - v1.0.0
Total Cycles: 1347

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2021 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 194.6 F 3.8
4:00 214.4 F 8.2
6:00 225.1 F 10.0
8:00 229.3 F 8.9
10:00 237.1 F 10.2
12:00 242.7 F 12.2
14:00 241.2 F 11.0
16:00 254.9 F 17.9
18:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.6 F 28.6
3:00 272.4 F 28.9
4:00 271.4 F 28.4

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 232.4 F 0.0
10:00 229.8 F 0.0
15:00 229.9 F 0.0
20:00 231.0 F 0.0
25:00 230.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

11/09/2021 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____