

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 001504  
15:17 10/14/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:37  
PRESSURIZING 6:15  
STERILIZING 7:44  
136.0 C 220kPa 7:44  
136.4 C 223kPa 8:14  
136.5 C 219kPa 8:44  
136.4 C 222kPa 9:14  
136.2 C 221kPa 9:44  
136.2 C 221kPa 10:14  
136.5 C 223kPa 10:44  
136.3 C 223kPa 11:14  
VENTING 11:14  
AIR DRYING 12:17  
CYCLE COMPLETE 21:54  
-----

Digital Signature #  
10AC87B269996145  
-----