
Midmark M11 - v1.0.5
Total Cycles: 72

Name: RITTER2
Sterilizer ID: V2313228
06/07/2021 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.5 F 0.0
2:00 173.1 F 3.5
4:00 188.7 F 9.2
6:00 221.2 F 7.2
8:00 228.3 F 11.3
10:00 235.8 F 9.7
12:00 236.5 F 9.1
14:00 250.9 F 16.2
16:00 262.8 F 23.4
18:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 271.8 F 29.7
2:00 271.9 F 29.8
3:00 271.9 F 29.8
4:00 271.6 F 29.5

Min 271.5 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 227.1 F 0.0
10:00 221.3 F 0.0
15:00 219.0 F 0.0
20:00 218.8 F 0.0
Dry Cycle Aborted

FILLING: 2:31
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 24:30
TOTAL CYCLE: 00:52:14

06/07/2021 01:52 PM
ACCEPTED: X
REJECTED:
Operator: U01