
Midmark M11 - v1.0.0
Total Cycles: 3186

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2023 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.6 F 0.0
2:00 180.1 F 1.1
4:00 195.5 F 5.0
6:00 215.2 F 10.1
8:00 223.5 F 6.8
10:00 235.0 F 11.2
12:00 234.1 F 8.6
14:00 239.8 F 10.3
16:00 246.7 F 13.9
18:00 259.3 F 20.7
20:00 270.4 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.4
2:00 272.1 F 28.8
3:00 271.6 F 28.7
4:00 272.8 F 29.5

Min 271.5 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 214.7 F 0.0
10:00 233.7 F 0.0
15:00 230.4 F 0.0
20:00 224.6 F 0.0
25:00 221.5 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

05/02/2023 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____