
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008130
15:45 02/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:34
STERILIZING 3:38
136.4 C 223kPa 3:38
136.4 C 225kPa 4:08
136.3 C 224kPa 4:38
136.4 C 225kPa 5:08
136.2 C 221kPa 5:38
136.4 C 225kPa 6:08
136.4 C 224kPa 6:38
136.0 C 221kPa 7:08
136.0 C 221kPa 7:08
VENTING 7:08
AIR DRYING 8:02
CYCLE COMPLETE 8:13

Digital Signature #
D79C7F398A21BD6B
